

## Test Report

Test bench serial number: 0183  
Transducer type: WIKA S-20  
Ordering number: -  
Range: 0-1000 bar  
Output signal: 4-20 mA  
Power supply: -  
Transducer serial number: 1A01DGO9FII  
Digital indicator: PLC+PC  
Line: B  
Station: 3

Test report n. **05\_10\_2024\_TR\_00000429**  
Customer: JC FABRICA DE VALVULAS S.A.U

### Certificate

**BIT02B - 924687 - LAT 093 170719P**

### Climbing measure

Pressure	Measure	Error limits measured (%)
0.00	0.00	0.00
175.00	175.00	0.00
350.00	350.00	0.00
525.00	525.00	0.00
700.00	701.00	0.10

### Downhill measure

Pressure	Measure	Error limits measured (%)
700.00	701.00	0.10
525.00	525.00	0.00
350.00	351.00	0.10
175.00	175.00	0.00
0.00	0.00	0.00



Date: 05/10/2024

Firma HPSC:

