

## Test Report

Test bench serial number: 0183  
 Transducer type: WIKA S-20  
 Ordering number: -  
 Range: 0-1000 bar  
 Output signal: 4-20 mA  
 Power supply: -  
 Transducer serial number: 1A01DGOEH20  
 Digital indicator: PLC+PC  
 Line: C  
 Station: 3

Test report n. **27\_04\_2025\_TR\_00000017**  
 Customer: JC FABRICA DE VALVULAS S.A.U

### Certificate

**BIT02B - 924687 - LAT 093 170719P**

### Climbing measure

| Pressure | Measure | Error limits measured (%) |
|----------|---------|---------------------------|
| 0.00     | 0.00    | 0.00                      |
| 175.00   | 175.00  | 0.00                      |
| 350.00   | 350.00  | 0.00                      |
| 525.00   | 525.00  | 0.00                      |
| 700.00   | 700.00  | 0.00                      |
|          |         |                           |
|          |         |                           |

### Downhill measure

| Pressure | Measure | Error limits measured (%) |
|----------|---------|---------------------------|
| 700.00   | 700.00  | 0.00                      |
| 525.00   | 525.00  | 0.00                      |
| 350.00   | 350.00  | 0.00                      |
| 175.00   | 175.00  | 0.00                      |
| 0.00     | 0.00    | 0.00                      |
|          |         |                           |
|          |         |                           |



Date: 27/04/2025

Firma HPSC:

