

## Test Report

Test bench serial number: 0183  
 Transducer type: WIKA S-20  
 Ordering number: -  
 Range: 0-1000 bar  
 Output signal: 4-20 mA  
 Power supply: -  
 Transducer serial number: 1A01DGO4724  
 Digital indicator: PLC+PC  
 Line: B  
 Station: 4

Test report n. **27\_04\_2025\_TR\_00000015**  
 Customer: JC FABRICA DE VALVULAS S.A.U

### Certificate

**BIT02B - 924687 - LAT 093 170719P**

### Climbing measure

Pressure	Measure	Error limits measured (%)
0.00	0.00	0.00
175.00	175.00	0.00
350.00	350.00	0.00
525.00	525.00	0.00
700.00	700.00	0.00

### Downhill measure

Pressure	Measure	Error limits measured (%)
700.00	700.00	0.00
525.00	525.00	0.00
350.00	350.00	0.00
175.00	175.00	0.00
0.00	0.00	0.00

Date: 27/04/2025



Firma HPSC:

