

Test Report

Test bench serial number: 0235
 Transducer type: WIKA S-20
 Ordering number: -
 Range: 0-1000 bar
 Output signal: 4-20 mA
 Power supply: -
 Transducer serial number: 1A017SQ7EIN
 Digital indicator: PLC+PC
 Line: A
 Station: -

Test report n. **27_04_2025_TR_00000005**
 Customer: JC FABRICA DE VALVULAS S.A.U

Certificate

BIT02B - 924687 - LAT 093 170719P

Climbing measure

Pressure	Measure	Error limits measured (%)
0.00	0.00	0.00
175.00	175.00	0.00
350.00	350.00	0.00
525.00	525.00	0.00
700.00	700.00	0.00

Downhill measure

Pressure	Measure	Error limits measured (%)
700.00	700.00	0.00
525.00	525.00	0.00
350.00	350.00	0.00
175.00	175.00	0.00
0.00	0.00	0.00

Date: 27/04/2025



Firma HPSC:

