

Test Report

Test bench serial number: 0235
 Transducer type: WIKA S-20
 Ordering number: -
 Range: 0-1000 bar
 Output signal: 4-20 mA
 Power supply: -
 Transducer serial number: 1A02KWMR137
 Digital indicator: PLC+PC
 Line: B
 Station: -

Test report n. **27_04_2025_TR_00000004**
 Customer: JC FABRICA DE VALVULAS S.A.U

Certificate

BIT02B - 924687 - LAT 093 170719P

Climbing measure

| Pressure | Measure | Error limits measured (%) |
|----------|---------|---------------------------|
| 0.00 | 0.00 | 0.00 |
| 175.00 | 175.00 | 0.00 |
| 350.00 | 350.00 | 0.00 |
| 525.00 | 525.00 | 0.00 |
| 700.00 | 700.00 | 0.00 |
| | | |
| | | |

Downhill measure

| Pressure | Measure | Error limits measured (%) |
|----------|---------|---------------------------|
| 700.00 | 700.00 | 0.00 |
| 525.00 | 525.00 | 0.00 |
| 350.00 | 350.00 | 0.00 |
| 175.00 | 175.00 | 0.00 |
| 0.00 | 0.00 | 0.00 |
| | | |
| | | |

Date: 27/04/2025



Firma HPSC:

